



3ply Mask

3ply Mask Detailed technical specifications



Product name	Disposable medical surgical masks
Specification	175 x 95 mm 3-ply flat ear type
Performance and composition	This product is composed of 3-ply (non-woven fabric + fusing fabric + non-woven fabric) hot-pressing mask, nose clip is made of plastic material and mask belt is made of non-woven fabric rubber cord.
Application	For all kinds of personnel to wear during non-invasive operations, for the prevention of directly through virus, bacteria and other pathogen microbial particles , to provide a physical barrier.
Validity	Fungus-free masks for two years, non - fungus masks for one year.

1 pack (contains 50 pcs)



50pcs per box



3ply mask CE certificate

شهادة – Certificat – 증명서 – 證明書 – Сертификат – Certificate – Certificat – 증명서 – 증명서

Form QAT_10-M04, version 00, effective since March 6th, 2020

Certificate of Compliance



No. 4S200406F.ADE0T91

Certificate's Holder: Anhui Dongguan Equipment Co., Ltd.
No. 688, Tianyang East road, Lulong Township,
Tianchang City, Anhui Province

Certification ECM Mark:

Product: Disposable face mask
Model(s): DG-QX01

Verification to: Standard: EN 149 : 2001+A1:2009
related to CE Directive(s): R 2016/425 (Personal Protective Equipment)

Remark: This document has been issued on a voluntary basis and upon request of the manufacturer. It is our opinion that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about its release and its use.

Additional information and clarification about the Marking:



The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a Notified Body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. RG01_ECM rev.3 available at: www.entecerma.it

Issuance date: 06 April 2020

Expiry date: 05 April 2025

Reviewer
Technical expert
Amanda Payne

Approver
ECM Service Director
Luca Bedonni

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it

3ply mask Test results

   中国认可
国际互认
检测
TESTING
CNAS L0606
171200110769

检 验 报 告

TEST REPORT

No: (2020) 皖检 XF 字 第 01301 号

产品名称 一次性防护型口罩
Product Name

受检单位 /
Inspected Body

委托单位 [REDACTED]
Client

检验类别 委托检验
Kind of Test

 安徽省产品质量监督检验研究院
Anhui Provincial Supervising & Testing Research Institute for Product Quality

安徽省产品质量监督检验研究院
检验报告附页

No: (2020) 皖检 XF 字 第 01301 号 共 2 页 第 2 页

序号	检验项目名称	技术要求	检验结果	单项判定	检测方法
1	甲醛含量 (mg/kg)	/	未检出	/	GB/T2912.1-2009
2	pH 值	/	7.2	/	GB/T7573-2009
3	断裂强力 (N) 口罩带 口罩带与口罩体的连接处	/	53 12	/	GB/T13773.2-2008
4	过滤效率 (%) 盐性介质	/	95.0	/	GB/T32610-2016

以下空白

 安徽省
 监督
用章

3ply mask Test results

安徽省产品质量监督检验研究院

Anhui Provincial Supervising & Testing Research Institute for Product Quality

检 验 报 告

TEST REPORT

No: (2020) 皖检 XF 字第 01301 号

共 2 页 第 1 页

产品名称 Product Name	一次性防护型口罩		型号规格 Model/Type	/	
标称生产单位 Nominal Manufacturer	/		受检单位 Inspected Body	/	
委托单位 Client	[Redacted]		抽样单位 Sampling Body	/	
委托单位地址 Client Address	/		抽样地点 Sampling Site	/	
检验项目 Test Items	共肆项【详见附页】		样品特性和状态 Sample Character and Condition	外观无异常	
检验日期 Test Date	2020.3.11—2020.3.14		原编号或生产日期 Serial Number/Manufactured Date	/	
标称商标/品牌 Nominal Trade Mark/Brand	/	抽样人员 Sealing Staff	/	检查封样人员 Checking and Sealing Samples	/
检验类别 Kind of Test	委托检验	抽样基数 Sampling Base	/	样品数量 Samples Quantity	40 只
样品等级 Sample Grade	/	抽样日期 Sampling Date	/	到样日期 Receipt Date	2020.03.11
检验依据 Test Criteria	GB/T32610-2016				
检验结论 Test Conclusion	该样品按 GB/T32610-2016 标准检验，检验结果见附页。				
备注 Note	此栏空白。				

质量
检验



批准:
Approved by:

黄成茹

审核:
Audited by:

王玮

主检:
Tested by:

唐思思

Boltpower Energy Group : Anhui Dongguan Equipment Co.ltd (letter of Authorization)

Ready as 24/4 morning time